The Arab American University Purchasing Procurement Dep.



الجامعة العربية الإمريكية دائرة اللوازم والمشتريات

| التاريخ: 14/05/2024 | رقم الاستدراج: | Q253.2024 |
|---|----------------|---------------|
| السادة: | | |
| تلفون: | | |
| إيميل: | | |
| استدراج عروض أسعار: الرجاء تزويدنا بأسعار اللوازم التالية شامل لضريبة القيمة المضافة: | | BR2324-000323 |

| اجمالي تكلفة البند | السعر | الكمية | الوحدة | اسم الصنف | # |
|--------------------|-------|---------------------|--------|---|---|
| | | 1 | عدد | Water Distiller(Small Volume) مرفق الوصف على الصفحات2.3 | 1 |
| | | 2 | عدد | Safety and Emergency Shower System with Eye & Face Wash Bowl, the price include all fittings needed | 2 |
| | دولار | المجموع الكلي (ش.ض) | | | |

| ختم وتوقيع المورد: | | تعليمات الاستدراج |
|--------------------|--------------------------------|--|
| | الثلاثاء | * آخر موعد لتسليم العروض هو يوم: |
| | 21/05/2024 | * الموافق: |
| | أجهزة مختبرات(تتعلق بالماء) | ** طبيعة البنود المطلوبة: |
| | دولار | * بعملة: |
| | الحرم الرئيسي - جنين | * الأسعار تشمل التوصيل ل: |
| | تورید کامل | * نوع التوريد: |
| | بالظرف المغلق | : * يرجى إرسال العرض ب: |
| | | شوط الاستدباح: |

- ** ضرورة تحديد مدة الكفالة إن وجدت.
- ** ضرورة تحديد مدة التوريد من تاريخ استلام امر الشراء، و في حال عدم الالتزام يتم فرض غرامة تأخير .
- ** ختم وتوقيع المورد المعتمد على هذا النموذج، وكتابة رقم الاستدراج واسم الشركة كما هو مسجل لدى الضريبة على الظرف.
 - ** للمراجعة والاستفسار بخصوص المواصفات يرجى الاتصال على هاتف 042418888 رقم داخلي 1481
- ** للمراجعة والاستفسار بخصوص المواصفات عن طريق فاكس 042510972 أو بريد الكتروني manal.daraghmah@aaup.edu
 - ** لمشاهدة العطاءات والاستدراجات، يرجى الدخول الى الرابط http://www.aaup.edu/tenders

مدير دائرة اللوازم و المشتريات أحسن ربايعة مدير على اللهائية على اللهائية المائية المائية المائية المائية المائية اللهائية المائية المائية المائية المائية المائية المائية ال

Water Distiller(Small Volume)

3years warranty, 10 years availability of spare parts.

Origin: Brand name USA, EURO, Japan

Water distiller with a capacity of ~6 l/hour.

- Ion exchange reverse osmosis system
- The product's source and manufacturing should be situated in the USA, Europe, or Japan, ensuring that both the design and production phases are conducted within these regions.
- · Easy to use and reliable at all times.
- Low cost, compared to the distiller
- · Bench top or under-bench designed.
- Distillate conductivity: $\sim 2.5 \mu \text{S/cm}$ (at 25° C)
- Daily average of 120 liters of water.
- A stainless steel tank designed for distilled water storage, possessing a capacity of at least 20 liters, or an equivalent alternative, would be suitable
- User and service manual in English.
- Please price all annual consumables for the device, otherwise, the price offer will not be considered.
- The system should be equipped with two TDS meters, one of which is integrated into the system for automatic monitoring, while the other is handheld and portable for added convenience.
 - السعر يشمل التوريد والتركيب ويجب أن يتم توضيح جميع متطلبات التركيب في عرض السعر من ملحقات ومتطلبات غير ذلك المورد ملزم بأية مصاريف أو أعمال يحتاجها الجهاز.
 - <u>Kindly your offer will not be taken into consideration if you don't fill in the table below.</u>

| tuble below. | | |
|---|--|---|
| Comment | Yes/No | Water Distiller(Small Volume) |
| | Manufacture | Manufacture. |
| | Model | Model. |
| | Origin | Origin: |
| | Weeks | Delivery time. |
| | Yes/No | 3 years Full warranty. |
| | Yes/No | Brand name. |
| | Yes/No | Water distiller with a capacity of ~6 l/hour. |
| | Yes/No | Ion exchange reverse osmosis system |
| | Yes/No | The product's source and manufacturing should be situated in the USA, Europe, or Japan, ensuring that both the design and production phases are conducted within these regions. |
| | Yes/No | Easy to use and reliable at all times. |
| The second section of the second | Water Water Company of the Company o | |



| Yes/No | Low cost, compared to the distiller |
|--------|---|
| Yes/No | Bench top or under-bench designed. |
| Yes/No | Distillate conductivity: ~ 2,5 μS/cm (at 25° C) |
| Yes/No | Daily average of 120 liters of water. |
| Yes/No | A stainless steel tank designed for distilled water storage, possessing a capacity of at least 20 liters, or an equivalent alternative, would be suitable |
| Yes/No | User and service manual in English. |
| Yes/No | Please price all annual consumables for the device, otherwise, the price offer will not be considered. |
| Yes/No | The system should be equipped with two TDS meters, one of which is integrated into the system for automatic monitoring, while the other is handheld and portable for added convenience. |
| Yes/No | User, Application, and service manuals. |

